

Nipro Medical Corporation

Issues a Voluntary Recall of All GlucoPro Insulin Syringes

Nipro Medical Corporation, Miami FL, is initiating a nationwide recall of all GlucoPro Insulin Syringes (This does not include the GlucoPro syringe specific for use with the Amigo Insulin pump). These syringes may have needles that detach from the syringe. If the needle becomes detached from the syringe during use, it can become stuck in the insulin vial, push back into the syringe, or remain in the skin after injection.

Consumers who have GlucoPro Insulin Syringes should stop using and return them to point of sale for reimbursement.

This recall includes all product codes and lot numbers with expiration dates before 2011-11 (Nov 1, 2011).

Product Code	NDC Number	Lot #	Expiry date
JD+01U3008-5C	41405-1308-01	A08022	2011-11
	41405-1308-10		
	41405-1308-50		
JD+01U3013-5C	41405-1303-01	A08013	2011-06
	41405-1303-10	A08017	2011-08
	41405-1303-50		
JD+01U3108-5C	41405-1318-01	A08013	2011-06
	41405-1318-10	A08017	2011-08
	41405-1318-50		
JD+03U3008-5C	41405-3308-01	C08022	2011-11
	41405-3308-10		
	41405-3308-50		
JD+03U3013-5C	41405-3303-01	C08013	2011-06
	41405-3303-10	C08017	2011-08
	41405-3303-50		
JD+03U3108-5C	41405-3318-01	C08013	2011-06
	41405-3318-10	C08017	2011-08
	41405-3318-50		
JD+05U3008-5C	41405-5308-01	B08022	2011-11
	41405-5308-10		
	41405-5308-50		
JD+05U3013-5C	41405-5303-01	B08013	2011-06
	41405-5303-10	B08017	2011-08
	41405-5303-50		
JD+05U3108-5C	41405-5318-01	B08013	2011-06
	41405-5318-10	B08017	2011-08
	41405-5318-50		

The firm voluntarily recalled the products after learning of the possibility of needle detachment. FDA has been apprised of this action.

No injuries have been reported to date.

Product was distributed nationwide, including Puerto Rico.

Company is notifying its distributors and customers by Fax and Email and is arranging for return of all recalled products.